

Resource Summary Report

Generated by [Kravitz](#) on Sep 10, 2026

Nikon: microscope camera DS-Fi3

RRID:SCR_018857

Type: Tool

Proper Citation

Nikon: microscope camera DS-Fi3 (RRID:SCR_018857)

Resource Information

URL: <https://www.microscope.healthcare.nikon.com/products/cameras/ds-fi3>

Proper Citation: Nikon: microscope camera DS-Fi3 (RRID:SCR_018857)

Description: High definition color microscope camera equipped with 5.9 megapixel CMOS image sensor which enables capture of high definition images of up to 2880 x 2048 pixels. Enables various observations such as brightfield, DIC, phase contrast and fluorescence.

Abbreviations: DS-Fi3

Resource Type: instrument resource

Keywords: microscope camera, CMOS image sensor, image capture, brightfield, phase contrast, fluorescence, instrument, equipment

Availability: Restricted

Resource Name: Nikon: microscope camera DS-Fi3

Resource ID: SCR_018857

Alternate URLs:

https://downloads.microscope.healthcare.nikon.com/phase4/literature/Brochures/DSseriesBio_2CE-MPAK-5.pdf

Record Creation Time: 20220129T080342+0000

Record Last Update: 20260905T132843+0000

Ratings and Alerts

No rating or validation information has been found for Nikon: microscope camera DS-Fi3.

No alerts have been found for Nikon: microscope camera DS-Fi3.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [Kravitz](#) .

Bagge J, et al. (2026) The histologic and molecular alterations during the cold storage of the ascending human colon. Cryobiology, 122, 105585.