

Resource Summary Report

Generated by T1D on Jul 31, 2026

Precisionary: Compresstome VF-700

RRID:SCR_018409

Type: Tool

Proper Citation

Precisionary: Compresstome VF-700 (RRID:SCR_018409)

Resource Information

URL: https://precisionary.com/wp-content/uploads/2018/06/Compresstome-VF-700-0Z-User-Manual_FINAL.pdf

Proper Citation: Precisionary: Compresstome VF-700 (RRID:SCR_018409)

Description: Vibrotome by Precisionary Instruments designed for cutting fixed tissue slices for histology studies. Offers two point stabilization of blade holder to eliminate unwanted chatter marks on surfaces of tissue slice. Fully automated to section fixed tissues with thickness range of 4 um to 2000 um with adjustable precision of 1 um.

Synonyms: Compresstome VF-700-0Z

Resource Type: instrument resource

Keywords: Precisionary Instrument, cutting fixed tissue, tissue slice, histology, two point stabilization blade holder, instrument, equipment

Availability: Restricted

Resource Name: Precisionary: Compresstome VF-700

Resource ID: SCR_018409

Alternate URLs: https://nbt ltd.com/wp-content/uploads/2019/11/compresstome-vf-700-0z-user-manual_final.pdf

Record Creation Time: 20220129T080340+0000

Record Last Update: 20260725T190848+0000

Ratings and Alerts

No rating or validation information has been found for Precisionary: Compresstome VF-700.

No alerts have been found for Precisionary: Compresstome VF-700.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [T1D](#) .

Benavidez NL, et al. (2021) Organization of the inputs and outputs of the mouse superior colliculus. Nature communications, 12(1), 4004.