

Resource Summary Report

Generated by [kravitz2](#) on Sep 3, 2026

Gilson: GX-271 HPLC System

RRID:SCR_018579

Type: Tool

Proper Citation

Gilson: GX-271 HPLC System (RRID:SCR_018579)

Resource Information

URL: <https://www.gilson.com/verity-271-hplc-system.html>

Proper Citation: Gilson: GX-271 HPLC System (RRID:SCR_018579)

Description: HPLC liquid handler with midsize bed capacity and fraction collector. It has maximum flow rates up to 200 mL per min, injection loop volumes of 2 uL to 25 mL, and max sample mass of 15 grams per injection.

Resource Type: instrument resource

Keywords: HPLC System, Instrument, Equipment, USEdit, Gilson, ABRF

Resource Name: Gilson: GX-271 HPLC System

Resource ID: SCR_018579

Alternate IDs: SCR_019963, Model_Number_GX_271

Record Creation Time: 20220129T080340+0000

Record Last Update: 20260829T182610+0000

Ratings and Alerts

No rating or validation information has been found for Gilson: GX-271 HPLC System.

No alerts have been found for Gilson: GX-271 HPLC System.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We have not found any literature mentions for this resource.