

Resource Summary Report

Generated by [PRECISE-TBI](#) on Sep 17, 2026

Eppendorf: Mastercycler nexus PCR Thermal Cycler

RRID:SCR_023266

Type: Tool

Proper Citation

Eppendorf: Mastercycler nexus PCR Thermal Cycler (RRID:SCR_023266)

Resource Information

URL: <https://www.eppendorf.com/gb-en/eShop-Products/PCR/Thermocyclers/Mastercycler-nexus-p-PF-14698>

Proper Citation: Eppendorf: Mastercycler nexus PCR Thermal Cycler (RRID:SCR_023266)

Description: Thermal cycler conduct PCR runs with range of standard vessel types, from 0.2 mL tubes to 96 well plates using single 96-well aluminum block. Can accommodate 96-well PCR plates, 0.2 mL PCR tubes, 0.2 mL PCR tube strips, and 0.5 mL PCR tubes. Mastercycler nexus cycler can email you when PCR run is complete.

Resource Type: instrument resource

Keywords: Eppendorf, Thermal cycler, PCR, 96-well PCR plates, 0.2 mL PCR tubes, 0.2 mL PCR tube strips, 0.5 mL PCR tubes, instrument, equipment, USEDI

Availability: Restricted

Resource Name: Eppendorf: Mastercycler nexus PCR Thermal Cycler

Resource ID: SCR_023266

Record Creation Time: 20230210T050214+0000

Record Last Update: 20260912T200037+0000

Ratings and Alerts

No rating or validation information has been found for Eppendorf: Mastercycler nexus PCR Thermal Cycler.

No alerts have been found for Eppendorf: Mastercycler nexus PCR Thermal Cycler.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [PRECISE-TBI](#) .

Hinsberger M, et al. (2026) Chromosome 1p and 6q Loss of Heterozygosity in Meningioma: A Comprehensive Analysis of the Two Chromatin Remodeling Complex Subunits ARID1A and ARID1B. Cancers, 18(9).