

Resource Summary Report

Generated by [PRECISE-TBI](#) on Aug 9, 2026

Stoelting: Touch Test Sensory Probes

RRID:SCR_025646

Type: Tool

Proper Citation

Stoelting: Touch Test Sensory Probes (RRID:SCR_025646)

Resource Information

URL: <https://stoeltingco.com/Neuroscience/Touch-Test-Sensory-Probes~9834>

Proper Citation: Stoelting: Touch Test Sensory Probes (RRID:SCR_025646)

Description: Semmes-Weinstein Von Frey Aesthesiometer for touch assessment. The Semmes-Weinstein monofilaments provide an approximate logarithmic scale of actual force, and a linear scale of perceived intensity. They have a long history of effective use in clinical settings, and can be used to diagnose pathologies of hyper- or hypo-aesthesia. Subsets within the set of 20 probes distinguish pathologies on different parts of the body (foot, hand, lip, cheek, etc.).

Resource Type: instrument resource

Keywords: Von Frey, Behavior assessment, Pain Assessment

Resource Name: Stoelting: Touch Test Sensory Probes

Resource ID: SCR_025646

Record Creation Time: 20240823T053238+0000

Record Last Update: 20260808T190728+0000

Ratings and Alerts

No rating or validation information has been found for Stoelting: Touch Test Sensory Probes.

No alerts have been found for Stoelting: Touch Test Sensory Probes.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [PRECISE-TBI](#) .

Wright A, et al. (2025) Repetitive Blast Exposure Drives Chronic Pain in Female Rats. Journal of neuroscience research, 103(12), e70103.