

Resource Summary Report

Generated by SPARC SAWG on Aug 9, 2026

WPI: Constant Current Stimulus Isolator

RRID:SCR_022660

Type: Tool

Proper Citation

WPI: Constant Current Stimulus Isolator (RRID:SCR_022660)

Resource Information

URL: <https://www.wpiinc.com/var-2273-constant-current-stimulus-isolator>

Proper Citation: WPI: Constant Current Stimulus Isolator (RRID:SCR_022660)

Description: Automated bipolar pulsing for zero net charge on biological preparations. Includes Constant current, Unipolar and bipolar stimulation modes, Built-in non-compliance alarm, Optically isolated input Standard TTL triggering, DC test mode. Powered by 9 V alkaline or rechargeable batteries.

Synonyms: Stimulus Isolator, WPI Constant Current Stimulus Isolator, Constant Current Stimulus Isolator

Resource Type: instrument resource

Keywords: Isolator, stimulus isolator, constant current isolator, constant current, World Precision Instruments, USEDit, instrument, equipment

Availability: Restricted

Resource Name: WPI: Constant Current Stimulus Isolator

Resource ID: SCR_022660

Alternate IDs: Model_Number_A365

Record Creation Time: 20220810T050149+0000

Record Last Update: 20260808T190235+0000

Ratings and Alerts

No rating or validation information has been found for WPI: Constant Current Stimulus Isolator.

No alerts have been found for WPI: Constant Current Stimulus Isolator.

Data and Source Information

Source: [SciCrunch Registry](#)

Usage and Citation Metrics

We found 1 mentions in open access literature.

Listed below are recent publications. The full list is available at [SPARC SAWG](#) .

Sinha AK, et al. (2026) Elucidating the Role of Muscarinic Acetylcholine Receptor Signalling in Efferent-Mediated Responses of Mammalian Vestibular Afferents. The European journal of neuroscience, 63(10), e70540.