

# Resource Summary Report

Generated by SPARC SAWG on Sep 17, 2026

## Nikon: microscope camera DS-Fi3

RRID:SCR\_018857

Type: Tool

### Proper Citation

Nikon: microscope camera DS-Fi3 (RRID:SCR\_018857)

### Resource Information

**URL:** <https://www.microscope.healthcare.nikon.com/products/cameras/ds-fi3>

**Proper Citation:** Nikon: microscope camera DS-Fi3 (RRID:SCR\_018857)

**Description:** High definition color microscope camera equipped with 5.9 megapixel CMOS image sensor which enables capture of high definition images of up to 2880 x 2048 pixels. Enables various observations such as brightfield, DIC, phase contrast and fluorescence.

**Abbreviations:** DS-Fi3

**Resource Type:** instrument resource

**Keywords:** microscope camera, CMOS image sensor, image capture, brightfield, phase contrast, fluorescence, instrument, equipment

**Availability:** Restricted

**Resource Name:** Nikon: microscope camera DS-Fi3

**Resource ID:** SCR\_018857

**Alternate URLs:**

[https://downloads.microscope.healthcare.nikon.com/phase4/literature/Brochures/DSseriesBio\\_2CE-MPAK-5.pdf](https://downloads.microscope.healthcare.nikon.com/phase4/literature/Brochures/DSseriesBio_2CE-MPAK-5.pdf)

**Record Creation Time:** 20220129T080342+0000

**Record Last Update:** 20260912T195906+0000

### Ratings and Alerts

No rating or validation information has been found for Nikon: microscope camera DS-Fi3.

No alerts have been found for Nikon: microscope camera DS-Fi3.

---

## Data and Source Information

**Source:** [SciCrunch Registry](#)

---

## Usage and Citation Metrics

We found 1 mentions in open access literature.

**Listed below are recent publications.** The full list is available at [SPARC SAWG](#) .

Bagge J, et al. (2026) The histologic and molecular alterations during the cold storage of the ascending human colon. Cryobiology, 122, 105585.